

COMMODITY DESCRIPTION

H.S.K No.	ITEM No.	DESCRIPTION	Unit	Quantity
		청력검사기 (Diagnostic Audiometer)	System	1

A. SPECIFICATION :

Clinical Audiometer (Main Body)

1 Set

- Two channel design
- Phones(TDH50P,high frequency)
- Bone & sound field
- Computer interface
- TEN,Quick SIN Test
- USB 단자를 통한 전송
- Air Frequency 125Hz to 20,000Hz
- Bone Frequency 250Hz to 8000Hz
- Sound Field Frequency 250Hz to 8000Hz
- Display 13.2 " or 8.4 "
- Intensity -10dB to 120dB

B. CONSISTING OF :

- | | |
|---|------|
| 1. Test Headset Assembly | 1 EA |
| 2. Bone Vibrator Assembly | 1 EA |
| 3. High Frequency Headset | 1 EA |
| 4. Monitor Headset(For Operator) | 1 EA |
| 5. Subject Response Swich | 1 EA |
| 6. Talkback Microphone(For Patient hearing) | 1 EA |
| 7. Power cord | 1 EA |
| 8. Service & Opeating Manual | 1 EA |

A. type

- Touch screen
- Talkback Microphone with mounting bracket
(operator speaking)

C. REMARKS :

1. Installation

계약자는 고객의 전제 조건으로 이 장비를 설치해야 하고, 제품 교육에 대해 무료로 제공됩니다.

2. Warranty

무상보증 조건. 3년의 무상 서비스 기간이 제공됩니다.

3. Data 전송을 위한 PACS 및 EMR 연동 구축비용 포함

* 위와 같이 일반 경쟁 입찰을 의뢰합니다.

PS : 입찰 후 최저가 응찰 업체를 낙찰자로 결정함.

2026. 06. 19.

사용부서명 : 이비인후과 청력검사실

신청인 성명 : 이이경 인이경

부서장 성명 : 김두연

기술부서명 : 의용공학과

담당자 성명 : 강 석 만 인

의용공학과장代 : 서 보 경 인